

### Continuous Radon Monitor

Model Number: 1028  
Calibration Date: 07/03/2008  
Monitor Time: 11/11/2008 17:17

Serial Number: 55540046  
CF: 2.84

780 Miller Drive, Oregon, WI 53575



### Inspector Information

Matthew Selvey  
Anytime Inspections L.L.C.  
780 Miller Drive  
Oregon  
WI-53575  
Phone Number: 608-358-3073  
License Number: 1340-106

### Customer Information

Matthew & Annette Selvey  
780 Miller Drive  
Oregon  
WI-53575  
Phone: 608 835 0511

780 Miller Drive, Oregon, WI 53575



### Site & Condition

Year Built: 1978  
SqFt: 1950


Structure Type: Two Story/Tudor  
Monitor Location: Unfinished Basement

### Test Summary

Start Time: 11/07/2008 12:38  
End Time: 11/10/2008 06:38  
Measurement Interval(hr): 1.0  
Elapse Time: 2 Days 18 hrs

Overall Avg: 2.8 pCi/l  
EPA Avg: 2.7 pCi/l

### Continuous Radon Monitor



Radon pCi/l     Motion Errors     Temp F     Rel % Humidity     Press inHg

\*\*\*\*\* 11/07/2008 \*\*\*\*\*

Time	Counts pCi/l	Flags
13:38	3.5	
14:38	3.9	
15:38	7.0	
16:38	5.3	
17:38	1.8	
18:38	4.2	
19:38	3.5	
20:38	2.8	
21:38	7.8	
22:38	3.9	
23:38	3.5	

\*\*\*\*\* 11/08/2008 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:38	5.6	
01:38	1.8	
02:38	3.9	
03:38	7.0	
04:38	6.7	
05:38	4.6	
06:38	4.6	
07:38	5.3	
08:38	2.1	
09:38	3.2	
10:38	6.7	
11:38	2.8	
12:38	2.1	
13:38	3.2	
14:38	2.1	
15:38	1.1	
16:38	2.1	
17:38	2.5	
18:38	3.2	
19:38	1.8	
20:38	0.7	
21:38	0.7	
22:38	2.5	
23:38	0.4	

\*\*\*\*\* 11/09/2008 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:38	1.1	
01:38	1.1	
02:38	1.1	

\*\*\*\*\* 11/09/2008 \*\*\*\*\*

Time	Counts pCi/l	Flags
03:38	2.8	
04:38	2.8	
05:38	2.8	
06:38	2.8	
07:38	0.4	
08:38	1.1	
09:38	1.4	
10:38	2.8	
11:38	1.8	
12:38	1.8	
13:38	1.4	
14:38	1.4	
15:38	3.2	
16:38	3.9	
17:38	3.2	
18:38	1.4	
19:38	1.4	
20:38	1.8	
21:38	2.8	
22:38	1.8	
23:38	2.1	

\*\*\*\*\* 11/10/2008 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:38	0.7	
01:38	1.8	
02:38	1.8	
03:38	1.8	
04:38	3.5	
05:38	2.5	
06:38	2.5	

Error Flags:

M Motion:

Inspector Signature \_\_\_\_\_

PC Software Version: 2.1.3.0  
Embedded Software Version: 106